

Communications protocol for the optimus sound level meters

Cirrus Research plc
Technical Note No. 48



The information contained within this document is ©Copyright Cirrus Research plc 2011.

All Rights Reserved.

All Trademarks Acknowledged.

Cirrus Research plc

Acoustic House

Bridlington Road

Hunmanby

North Yorkshire

YO14 0PH

United Kingdom

Tel: 0845 230 2434 (UK)

Tel: +44 1723 891655 (International)

Fax: +44 1723 891742

Email: sales@cirrusresearch.co.uk

Web: www.cirrusresearch.co.uk

Twitter: @cirrusresearch

Optimus® is a registered trademark of Cirrus Research plc. Acoustic Fingerprint™ and AuditStore™ trademarks pending.

Version 1.1

19th August 2013

1	Revision History	4
1.1	Version 1.1	4
2	Command Summary	5
3	Basic Information.....	6
3.1	Connecting cable	6
3.2	Data rate & parity.....	6
3.3	Command & response termination.....	6
3.4	Command case	6
3.5	Un-recognised commands	6
3.6	Command language.....	6
3.7	Command echo.....	6
3.8	Instrument settings	6
4	Baud Rate Settings	7
5	Help.....	7
6	Instrument Identification.....	8
7	Instrument Detailed Information.....	8
8	Clock	9
8.1	Setting the clock	9
9	Measuring.....	10
9.1	Check current measuring status.....	10
9.2	Start a new measurement.....	10
9.3	Stop the measurement	10
9.4	Reset the measurement	10
10	Live Data.....	11
10.1	Check current status.....	11
10.2	Start sending live data.....	11
10.3	Stop sending data	11
10.4	Request a one-off set of live data	12
10.4.1	1:1 Octave Bands	13
10.4.2	1:3 Octave Bands	14
11	Previous Measurement Data	15
12	Reboot	15
13	Examples of data communication	16
13.1	Instrument connection & live data.....	16
13.2	Instrument connection, start measurement & get measurement results	16
13.3	Repeating measurements	16
14	Appendix 1 Multi-IO RS232 Connector.....	18
14.1	Connector part numbers.....	18
15	Appendix 2 ZL:175 Optimus RS232 to PC Cable.....	19
16	Cirrus Research Offices	20

1 Revision History

1.1 Version 1.1

Additions

- Section 2 Command Summary
- Section 3.8 Instrument settings
- Section 13 Examples of data communications

Amendments

- Section 5 Help
- Section 10 Live Data
- Document formatting

Introduction

This document describes the RS232 serial command protocol for the Cirrus Research optimus sound level meters.

This document is subject to change without notice. The latest version of this document can be downloaded from the Cirrus Research plc website at www.cirrusresearch.co.uk/library

2 Command Summary

Command	Summary	See page
HELP?	Displays basic information along with a list of available commands	7
IDN?	Displays the basic identification information from the instrument	8
INFO?	Displays additional information available from the instrument	8
INFO2?	Displays additional information available from the instrument	8
CLOCK?	Displays the current time & date	9
CLOCK	Sets the time & date	9
MEASURE?	Checks the current measurement status	10
MEASURE START	Starts a new measurement	10
MEASURE STOP	Stops the current measurement	10
MEASURE RESET	Resets the current measurement	10
LIVE?	Displays the current live data status	11
LIVE START	Starts the transmission of live data	11
LIVE STOP	Stops the transmission of live data	11
LIVE NOW	Requests a single set of live data	12
PREV	Request a one-off set of data for the most recent completed measurement	15
REBOOT	Reboots the instrument	15

3 Basic Information

3.1 Connecting cable

Communication over RS232 with the optimus requires the ZL:175 Optimus RS232 to PC Cable

3.2 Data rate & parity

The baud rate is 115200 (default) or 9600 baud.
See 'Baud Rate Settings' below for details

Data Bits:	8
Parity:	None
Stop Bits:	1
Flow Control:	None

3.3 Command & response termination

All commands are terminated by CRLF ('\r\n').
All responses are terminated by CRLF ('\r\n').

3.4 Command case

Commands and Responses are not case sensitive but are defined here in uppercase.

3.5 Un-recognised commands

Unrecognised commands are ignored.

3.6 Command language

All commands are in English.

3.7 Command echo

Commands are not echoed back.

3.8 Instrument settings

All instrument settings for the optimus sound level meters are available via the NoiseTools software.

Measurement timers and Ln values are configured via the NoiseTools software.

Single or repeating measurement timers can be configured from 1 minute upwards via the NoiseTools software.

4 Baud Rate Settings

The Optimus has two choices of baud rate: 115200 baud (default) and 9600 baud. The baud rate is set using NoiseTools.

In the 'Advanced' section of the 'Configure' window in NoiseTools, select the 'Port A' setting as follows:

- 'Terminal' – 115200 baud
- 'SlowTerminal' – 9600 baud

5 Help

This command gives some basic information along with a list of available commands.

Command	HELP?
Response	<Basic connection information> <List available of commands, 1 per line>

Example

TX: HELP?
RX: Optimus Communications Protocol: RX: Commands are case insensitive, terminated with CRLF. RX: Responses are terminated with CRLF. RX: IDN? RX: INFO? RX: LIVE? <... all commands ...>

Notes:

- This is intended to aid development or for manual use. It serves no purpose for automated use as part of the actual communications protocol.

6 Instrument Identification

This command displays the basic identification information from the instrument.

Command	IDN?
Response	IDN <type> <serial> <version>

<type> is the type, or model number, of the instrument.

<serial> is the unique serial number of the instrument.

<version> is the current firmware version.

Example

TX: IDN?
RX: IDN CR:123A G123456 1.0.1234

Notes:

- Any spaces within the type and serial strings are converted to '_'.

7 Instrument Detailed Information

This command displays additional information available from the instrument.

Command	INFO?
Response	INFO <info list>

<info list> is a series of space-separated integers as below:

- Battery Voltage
 - 0 mV
- Reserved
- Preamp Temperature
 - 0 Degrees Kelvin - Subtract 273 for Celsius
- Reserved
- Reserved
- Reserved
- Reserved
- Reserved
- Reserved
- Date and Time
 - 0 Seconds since 01/01/1970 00:00:00

This command displays the 2nd additional information available from the instrument.

Command	INFO2?
Response	INFO2 <info2 list>

<info2 list> is a series of space-separated floats as below:

- Number of Measurements.

- Measurement Memory Free
 - o Percentage
- Audio Memory Free
 - o Percentage
- Modem Signal Strength
 - o Range 0 to 31. Low to High
 - o 99 = Unknown
- Modem Connection Data Type
 - o 0 = GPRS
 - o 1 = EGPRS
 - o 2 = WCDMA
 - o 3 = UMTS
 - o 4 = Unknown
- GPS Dilution of precision (hdop)
- GPS latitude
- GPS longitude

Examples

TX: INFO?
RX: INFO 5426 0 293 0 0 0 0 0 0 1370427877
TX: INFO2?
RX: INFO2 40 98.54 94.88 99 4 0.0 0.00000000 0.00000000

8 Clock

Checking the current time

Command	CLOCK?
Response	CLOCK YYYY-MM-DDTHH:MM:SS

8.1 Setting the clock

Command	CLOCK YYYY-MM-DDTHH:MM:SS
Response	CLOCK YYYY-MM-DDTHH:MM:SS or CLOCK FAILED

Example

TX: CLOCK?
RX: CLOCK 2010-01-01T15:30:00

Notes:

- All times use the ISO 8601 format, although without time zone.

9 Measuring

9.1 Check current measuring status

Command	MEASURE?
Response	MEASURE RUNNING or MEASURE STOPPED

9.2 Start a new measurement

Command	MEASURE START
Response	MEASURE RUNNING

Notes:

- MEASURE START skips the voice tag and starts recording immediately.
- If the single or repeat timers are enabled in the instrument, these will be activated by the MEASURE START command.
- Live data is available every 1 second when the instrument is measuring

9.3 Stop the measurement

Command	MEASURE STOP
Response	MEASURE STOPPED

Notes:

- MEASURE STOP will over-ride single or repeat timers.

9.4 Reset the measurement

Command	MEASURE RESET
Response	MEASURE RUNNING or MEASURE STOPPED

Measurement timers

10 Live Data

Please refer to the end of Live Data section for details of <list>, <data> and <extra> parameters.

The live data values are available once per second.

10.1 Check current status

Command	LIVE?
Response	LIVE RUNNING <list> or LIVE STOPPED

Example

TX: LIVE?
RX: LIVE RUNNING LAI

10.2 Start sending live data

One response will be returned per second until it is stopped.

Command	LIVE START <list>
Response	LIVE RUNNING <list> LIVE <data> <duration> <extra>

Example

TX: LIVE START LAEQ LAF
RX: LIVE RUNNING LAF LAEQ
RX: LIVE 9.73 10.27 2300.000 FFT
RX: LIVE 10.36 10.62 2301.000 FFT
RX: LIVE 10.12 10.62 2302.000 FFT

10.3 Stop sending data

Command	LIVE STOP
Response	LIVE STOPPED

10.4 Request a one-off set of live data

Command	LIVE NOW <list>
Response	LIVE NOW <list> LIVE <data> <duration> <extra>

<list> contains a list of data types.

<data> contains a list of values.

<duration> the run duration in seconds (to 3 decimal places).

<extra> contains state information which is always sent at the end of the list of values.

- Overload (1s)
 - T or F. (True or False)
- Measurement Overload
 - T or F. (True or False)
- Measurement Running
 - T or F. (True or False)

List of available data types.

Sound Pressure Level (SPL)	
Lxy	Instantaneous
x is the Frequency Weighting (A, C or Z) y is the Time Weighting (F, S, or I)	
LxyMAXT	Overall
x is the Frequency Weighting (A, C or Z) y is the Time Weighting (F, S, or I)	
LxyMINT	Overall
x is the Frequency Weighting (A, C or Z) y is the Time Weighting (F, S, or I)	
Integrated Level (Leq)	
LxEQ	Most recent 1s
x is the Frequency Weighting (A, C or Z)	
LEQ2	Most recent 1s
LEQ3	Most recent 1s
LxEQT	
x is the Frequency Weighting (A, C or Z)	
LEQ2T	Overall
LEQ3T	Overall
Peak Level	
LxPEAK	Instantaneous
x is the Frequency Weighting (A, C or Z)	
LxPEAKT	Overall
x is the Frequency Weighting (A, C or Z)	
Octave Integrated Level	
OCTz	Most recent 1s
z is the Band Number (0 to 9) See table below.	

OCTzT	Overall
z is the Band Number (0 to 9) See table below.	
Third Octave Integrated Level	
3OCTz	Most recent 1s
z is the Band Number (0 to 35) See table below.	
3OCTzT	Overall
z is the Band Number (0 to 35) See table below.	
Statistical Levels	
LNz	Overall
z is the Ln Percentage (1, 5, 10, 50, 90, 95 or 99)	
USERLNz	Overall
z is the User Ln Number (1 to 7) Configured through NoiseTools.	

Notes:

- A data type with the suffix 'T' denotes an overall value.
 - When a measurement is in progress, it is the cumulative value over the current run duration.
 - When not measuring, it is the final value of the most recently completed measurement.
- A data type without this suffix is the current value, which can be:
 - Based on the most recent 1s. Example: LAEQ
 - An instantaneous value. Example: LAF
- Some instruments do not support all the values listed above.
 - Any unsupported values are ignored when requested and will not be in the returned list.
- Overall values requested before a measurement is started, are shown as 'NaN'. Alternatively if a measurement has been made previously those values will be shown.
 - Instruments without storage only support overall values while running.
- All decibel values are rounded to two decimal places.
- The data is in the order specified by the list returned from the LIVE START or LIVE NOW commands, not the order in which they were requested.

10.4.1 1:1 Octave Bands

Band Number	Frequency (Hz)
0	31
1	62.5
2	125
3	250
4	500
5	1000
6	2000
7	4000
8	8000
9	16000

10.4.2 1:3 Octave Bands

Band Number	Frequency (Hz)	Band Number	Frequency (Hz)	Band Number	Frequency (Hz)
0	6.3	12	100	24	1600
1	8	13	125	25	2000
2	10	14	160	26	2500
3	12.5	15	200	27	3150
4	16	16	250	28	4000
5	20	17	315	29	5000
6	25	18	400	30	6300
7	31.5	19	500	31	8000
8	40	20	625	32	10000
9	50	21	800	33	12500
10	62.5	22	1000	34	16000
11	80	23	1250	35	20000

Examples

TX: LIVE?
RX: LIVE STOPPED
TX: LIVE START LAEQT LAF LAEQ LCPEAKT
RX: LIVE RUNNING LAF LAEQ LAEQT LCPEAKT
RX: LIVE 50.31 65.81 60.17 53.97 0.000 FFF
RX: LIVE 50.32 65.82 60.17 53.96 0.000 FFF
RX: LIVE 50.33 65.83 60.15 53.95 0.000 FFF
RX: LIVE 50.34 65.84 60.13 53.91 0.000 FFF
TX: LIVE?
RX: LIVE RUNNING LAF LAEQ LAEQT LCPEAKT
TX: LIVE START LAF
RX: LIVE RUNNING LAF
RX: LIVE 65.81 5.016 FFT
RX: LIVE 65.82 6.016 FFT
RX: LIVE 65.83 7.016 FFT
TX: LIVE STOP
RX: LIVE STOPPED
TX: LIVE NOW LAEQT LAEQ
RX: LIVE NOW LAEQ LAEQT
RX: LIVE 50.35 60.16 17.500 FFT

11 Previous Measurement Data

Request a one-off set of data for the most recent completed measurement.

Command	PREV <list>
Response	PREV <list> PREV <data> <datetime> <extra>

<list> contains a list of data types.

<data> contains a list of values.

<datetime> is in the same format as the CLOCK command.

<extra> is a single flag for the latched measurement overload.

Notes:

- The requested data types should only use overall values (with suffix T). Items without T will be return as 'NaN'.
- This command is only supported on instruments with the storage capability.

Example

TX: PREV LAEQT LCEQT LCPEAKT LZPEAKT
RX: PREV LAEQT LCEQT LCPEAKT LZPEAKT
RX: PREV 47.91 56.92 89.15 89.18 2011-09-23T12:25:02 F

12 Reboot

Reboot the instrument

Command	REBOOT
Response	REBOOT

Notes:

- The Instrument will wait 5 seconds before rebooting.

13 Examples of data communication

13.1 Instrument connection & live data

This example shows how to connect to an instrument and enable live data for monitoring purposes.

TX: IDN?
RX: IDN CR:171B G786430 2.5.1839
TX: LIVE START LAF LAS LAEQ LCEQ LCPEAK
RX: LIVE RUNNING LAF LAS LAEQ LCEQ LCPEAK
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
RX: LIVE 84.50 83.20 64.35 67.59 87.84 0.000 FFF
...

13.2 Instrument connection, start measurement & get measurement results

This example shows how to connect to an instrument, make a measurement and get the results.

TX: IDN?
RX: IDN CR:171B G786430 2.5.1839
TX: MEASURE START
RX: MEASURE RUNNING
<i>Wait 5 minutes</i>
TX: MEASURE STOP
RX: MEASURE STOPPED
TX: LIVE NOW LAFMAXT LASMAXT LAEQT LCEQT LCPEAKT LN90
RX: LIVE NOW LAFMAXT LASMAXT LAEQT LCEQT LCPEAKT LN90
RX: LIVE 85.50 84.20 65.35 68.59 88.84 42.50 300.000 FFF

When the instrument is stopped, the T commands (LAFMAXT, LASMAXT etc) return the data for the last measurement stored.

13.3 Repeating measurements

It is possible to configure the instrument via NoiseTools to use either the single timer or repeat timer to make measurements of the correct duration. If using the repeat timer see the multiple measurement example below.

This example shows how to connect to an instrument, start making a series of measurements and get the results.

TX: IDN?
RX: IDN CR:171B G786430 2.5.1839
TX: MEASURE START
RX: MEASURE RUNNING
<i>Wait 5 minutes</i>
TX: MEASURE STOP
RX: MEASURE STOPPED
TX: MEASURE START

RX: MEASURE RUNNING
TX: PREV LAFMAXT LASMAXT LAEQT LCEQT LCPEAKT LN90
RX: PREV LAFMAXT LASMAXT LAEQT LCEQT LCPEAKT LN90 RX: PREV 85.50 84.20 65.35 68.59 88.84 42.50 2011-09- 23T12:25:02 F
<i>Wait 5 minutes</i> <i>Repeat (STOP, START, PREV)</i>

If using repeat timers you do not need to run the START, STOP commands, simply wait the required amount of time then request the values with PREV.

14 Appendix 1 Multi-IO RS232 Connector

The 18 pin Multi-IO connector on the optimus sound level meter provides access to the RS232 data signal. This connector should be used with the ZL:175 Optimus RS232 to PC Cable (see Appendix 2).

The connections for the 18 way socket are as follows:



Pin	Colour	Description
1		Not connected
2		Not connected
3	Black	Ext_On
4	Turquoise	AC Out
5	Grey	DC Out
6		Not connected
7	Green	DC Out
8	White	TX0
9	Yellow	CTS0
10	Blue	RX0
11	Brown	ACT1
12	Orange	CTS1
13	Pink	TX1
14		Not connected
15		Not connected
16	Red	External Power(12Vdc)
17	Purple	RX1
18	Braid	Ground

Note: BRAID to be connected to metal shield.

14.1 Connector part numbers

Hirose ST40 18-WAY PLUG - ST40X-18S-CVR (with shield)

Hirose - ST40X-CM(4.0-5.0) - CLAMP BRACKET



Rear View

15 Appendix 2 ZL:175 Optimus RS232 to PC Cable

16 Cirrus Research Offices

The addresses given below are the Cirrus Research plc offices. Cirrus Research plc also has approved distributors in many countries worldwide. For details of your local representative, please contact Cirrus Research plc at the address below. Contact details for Cirrus Research authorised distributors and agents are also available from the Internet Web site at the address shown below.

Main Office

Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
United Kingdom
YO14 0PH

Telephone: +44 (0)1723 891655
Fax: +44 (0)1723 891742
E-mail: sales@cirrusresearch.co.uk
Web Site: www.cirrusresearch.co.uk

Germany

Cirrus Research plc Deutschland
Arabella Center
Lyoner Strasse 44 – 48
D-60528 Frankfurt
Germany

Tel: +49 (0)69 95932047
Fax: +49 (0)69 95932049
E-mail: vertrieb@cirrusresearch.de
Website: www.cirrusresearch.de

Spain

Cirrus Research SL
Travesera de Gràcia, 62 4o 7a
Barcelona
España
Teléfono: (+34) 93 362 28 91
E-mail: info@cirrusresearch.es
Web: www.cirrusresearch.es

France

Cirrus Recherche Ltd
40 Bis Avenue Gabriel Fauré
09500 Mirepoix
France

Tel: +33 5 61 67 40 01
Fax: +33 5 61 67 40 56
Email: sales@cirrusresearch.fr
Web: www.cirrusresearch.fr

Cirrus Environmental

Unit 2 Bridlington Road Industrial Estate
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Tel: +44 (0) 1723 891722
Email: sales@cirrus-environmental.com
Web: www.cirrus-environmental.com